A I		0	- : : : -	- 4: -	
Anal	yτe	Spe	CITIC	atic	ns:

IL-5

**Catalog Number:** 


92062

MultiCyt® QBeads™ PlexScreen Mouse Secreted Protein

Reconstituting the IL-5 standard in 2 mL of assay diluent results in a concentration of 5,000 pg/mL. Standard curves for no-wash and 1-wash protocols are shown starting from 5,000 pg/mL and proceeding in a 1:2 serial titration. The limit of detection, represented by the dashed line, is determined by adding 3 standard deviations to the mean of samples without IL-5 (N=12). Results are typical and may vary based on specific assay conditions.

## **No-Wash Assay Performance**

## 1-Wash Assay Performance



Lowest Detectable Concentration Tested (pg/mL): 20

Lowest Detectable Concentration Tested (pg/mL): 2.4

For research purposes only. Not for resale.

Rev. 1 Sep 2014